

APPENDIX 11.2

EQUIPMENT CALIBRATION CERTIFICATES



NSAI

National Metrology Laboratory

Certificate of Calibration

Issued to Fehily Timoney & Company
J5 Plaza
North Business Park
North Road
Dublin 11

Attention of John Mahon

Certificate Number	180300
Item Calibrated	Svantek SVAN 977 Sound Level Meter with ACO 7052E Microphone
Serial Number	34876 (SLM) and 56429 (Microphone)
Client ID Number	#2
Order Number	6252
Date Received	24 Jan 2018
NML Procedure Number	AP-NM-09

Method The above sound level meter was allowed to stabilise for a suitable period in laboratory conditions. It was then calibrated by carrying out the verification tests detailed in IEC 61672-3 (2006), *Periodic tests, specification for the verification of sound level meters*. This standard specifies a procedure for the periodic verification of conformance of a sound level meter or integrating-averaging meter to IEC 61672-1 (2003).

Calibration Standards Norsonic 1504A Calibration System incorporating:
SR DS360 Signal Generator, No. 0735 [Cal Due Date: 21 Dec 2018]
Agilent 34401A Digital Multimeter, No. 0736 [Cal Due Date: 17 Nov 2018]
B&K 4134 Measuring Microphone, No. 0743 [Cal Due Date: 28 Apr 2019]
B&K 4228 Pistonphone, No. 0740 [Cal Due Date: 21 Mar 2019]
B&K 4226 Acoustical Calibrator, No. 0150 [Cal Due Date: 15 May 2018]

Calibrated by


David Fleming

Approved by


Paul Hetherington

Date of Calibration

31 Jan 2018

Date of Issue

31 Jan 2018



This certificate is consistent with Calibration and Measurement Capabilities (CMC's) that are included in Appendix C of the Mutual Recognition Arrangement (MRA) drawn up by the International Committee for Weights and Measures. Under the MRA, all participating institutes recognize the validity of each other's calibration certificates and measurement reports for quantities, ranges and measurement uncertainties specified in Appendix C (for details see www.bipm.org)



NSAI

National Metrology Laboratory

Certificate of Calibration

Issued to

Fehily Timoney & Company
J5 Plaza
North Park Business Park
North Road
Dublin 11

Certificate Number

242697

Item Calibrated

Svantek SVAN 977 Sound Level Meter with ACO 7052E Microphone

Serial Number

69552 (SLM) and 69543 (Microphone)

ID Number

None

Order Number

7629

Date Received

10 Jun 2024

NML Procedure Number

TFAP-NM-16

Method

The above sound level meter was allowed to stabilise for a suitable period in laboratory conditions. It was then calibrated by carrying out the verification tests detailed in IEC 61672-3 (2006), *Periodic tests, specification for the verification of sound level meters*. This standard specifies a procedure for the periodic verification of conformance of a sound level meter or integrating-averaging meter to IEC 61672-1 (2003).

Calibration Standards

Norsonic 1504A Calibration System incorporating:
SR DS360 Signal Generator, No. 0735 [Cal Due Date: 01 Sep 2024]
Agilent 34401A Digital Multimeter, No. 0736 [Cal Due Date: 01 Sep 2024]
B&K 4180 Measuring Microphone, No. 1069 [Cal Due Date: 15 Sep 2025]
B&K 4228 Pistonphone, No. 0741 [Cal Due Date: 14 Sep 2025]
B&K 4226 Acoustical Calibrator, No. 0150 [Cal Due Date: 04 Jan 2025]

Calibrated by

David Fleming

Approved by

Rory Hanrahan

Date of Calibration

03 Jul 2024

Date of Issue

03 Jul 2024



This certificate is consistent with Calibration and Measurement Capabilities (CMC's) that are included in Appendix C of the Mutual Recognition Arrangement (MRA) drawn up by the International Committee for Weights and Measures. Under the MRA, all participating institutes recognize the validity of each other's calibration certificates and measurement reports for quantities, ranges and measurement uncertainties specified in Appendix C (for details see www.bipm.org)



NSAI

National Metrology Laboratory

Certificate of Calibration

Issued to Fehily Timoney & Company
J5 Plaza
North Park Business Park
North Road
Dublin 11

Attention of John Mahon

Certificate Number	173660
Item Calibrated	Svantek SVAN 977 Sound Level Meter with ACO 7052E Microphone
Serial Number	34173 (SLM) and 54691 (Microphone)
Client ID Number	#3
Order Number	6308
Date Received	20 Oct 2017
NML Procedure Number	AP-NM-09

Method The above sound level meter was allowed to stabilise for a suitable period in laboratory conditions. It was then calibrated by carrying out the verification tests detailed in IEC 61672-3 (2006), *Periodic tests, specification for the verification of sound level meters*. This standard specifies a procedure for the periodic verification of conformance of a sound level meter or integrating-averaging meter to IEC 61672-1 (2003).

Calibration Standards Norsonic 1504A Calibration System incorporating:
SR DS360 Signal Generator, No. 0735 [Cal Due Date: 18 Nov 2017]
Agilent 34401A Digital Multimeter, No. 0736 [Cal Due Date: 07 Nov 2017]
B&K 4134 Measuring Microphone, No. 0742 [Cal Due Date: 18 Jan 2018]
B&K 4228 Pistonphone, No. 0741 [Cal Due Date: 08 Jan 2018]
B&K 4226 Acoustical Calibrator, No. 0150 [Cal Due Date: 15 May 2018]

Calibrated by


David Fleming

Approved by


Paul Hetherington

Date of Calibration

31 Oct 2017

Date of Issue

01 Nov 2017



This certificate is consistent with Calibration and Measurement Capabilities (CMC's) that are included in Appendix C of the Mutual Recognition Arrangement (MRA) drawn up by the International Committee for Weights and Measures. Under the MRA, all participating institutes recognize the validity of each other's calibration certificates and measurement reports for quantities, ranges and measurement uncertainties specified in Appendix C (for details see www.bipm.org)

ENVIRONMENTAL CONDITIONS

Temperature	Relative humidity	Ambient pressure
29 °C	37%	1005 hPa

TEST EQUIPMENT

Item	Manufacturer	Model	Serial no.	Description
1.	SVANTEK	SVAN 401	127	Signal generator
2.	SVANTEK	SVAN 912A	9537	Sound & Vibration Analyser
3.	KEITHLEY	2000	0910165	Digital multimeter
4.	SVANTEK	SV33	48878	Acoustic calibrator
5.	SVANTEK	ST02	-	Microphone equivalent electrical impedance (18pF)
6.	DYTRAN	3333A	1376	Reference accelerometer

CONFORMITY & TEST DECLARATION

- Herewith Svantek company declares that this instrument has been calibrated and tested in compliance with the internal ISO9001 procedures and meets all specification given in the Manual(s) or respectively surpass them.
- The acoustic calibration was performed using the Sound Calibrator and is traceable to the GUM (Central Office of Measures) reference standard - sound level calibrator type 4231 No 2292773.
- The vibrational calibration was performed using the Back-to-Back Comparison method and is traceable to the GUM (Central Office of Measures) reference standard - accelerometer type 8305 No 1435233.
- The information appearing on this sheet has been compiled specifically for this instrument. This form is produced with advanced equipment & procedures which permit comprehensive quality assurance verification of all data supplied herein.
- This calibration sheet shall not be reproduced except in full, without written permission of the SVANTEK Ltd.

Calibration specialist: Krzysztof Kubel



Test date: 2018-06-01



ISO9001 certified

FACTORY CALIBRATION DATA OF THE SVAN 977A No. 69557

with preamplifier SVANTEK type SV12L No. 72140 and microphone ACO PACIFIC type 7052E No. 69617

SOUND LEVEL METER

1. CALIBRATION (electrical)

LEVEL METER function; Characteristic: A; $f_{\text{ref}}=1$ kHz; Input signal = 110.9 dB;

Range	Low (120dB)	High (137dB)
Indication [dB]	114.0	114.0
Error [dB]	0.0	0.0

2. CALIBRATION* (acoustical)

LEVEL METER function; Range: High; Reference frequency: 1000 Hz; Sound Pressure Level: 113.99 dB.

Characteristic	Correct value [dB]	Indication [dB]	Error [dB]
Z	113.99	113.97	-0.02
A	113.99	113.97	-0.02
C	113.99	113.97	-0.02

Calibration measured with the microphone ACO PACIFIC type 7052E No. 69617. Calibration factor: 0.02 dB.

3. LINEARITY TEST* (electrical)

LEVEL METER function; Range: Low; Characteristic: A; $f_{\text{ref}}=31.5$ Hz

Nominal result LEQ [dB]	24.0	25.0	26.0	28.0	30.0	40.0	60.0	80.0
Error [dB]	0.1	0.1	0.0	0.0	0.0	0.0	0.0	0.0

LEVEL METER function; Range: Low; Characteristic: A; $f_{\text{ref}}=1000$ Hz

Nominal result LEQ [dB]	24.0	25.0	26.0	28.0	30.0	40.0	60.0	80.0	100.0	120.0
Error [dB]	0.1	0.0	0.0	-0.0	-0.0	-0.0	-0.0	-0.0	-0.0	-0.0

LEVEL METER function; Range: Low; Characteristic: A; $f_{\text{ref}}=8000$ Hz

Nominal result LEQ [dB]	24.0	25.0	26.0	28.0	30.0	40.0	60.0	80.0	100.0	119.0
Error [dB]	0.0	-0.0	-0.0	-0.0	-0.0	-0.0	-0.0	-0.0	-0.0	-0.0

LEVEL METER function; Range: High; Characteristic: A; $f_{\text{ref}}=31.5$ Hz

Nominal result LEQ [dB]	35.0	36.0	37.0	38.0	40.0	60.0	80.0	97.0
Error [dB]	0.1	0.1	0.1	0.0	0.0	-0.0	-0.0	0.0

LEVEL METER function; Range: High; Characteristic: A; $f_{\text{ref}}=1000$ Hz

Nominal result LEQ [dB]	35.0	36.0	37.0	38.0	40.0	60.0	80.0	100.0	120.0	137.0
Error [dB]	0.1	0.1	0.0	0.0	0.0	-0.0	-0.0	0.0	-0.0	-0.0

LEVEL METER function; Range: High; Characteristic: A; $f_{\text{ref}}=8000$ Hz

Nominal result LEQ [dB]	35.0	36.0	37.0	38.0	40.0	60.0	80.0	100.0	120.0	136.0
Error [dB]	0.0	0.0	0.0	-0.0	0.0	-0.0	-0.0	-0.0	-0.0	0.0

4. TONE BURST RESPONSE*

LEVEL METER function; Characteristic: A; $f_{\text{ref}}=4000$ Hz; Burst duration: 2 s

Range: Low; Steady level nominal result = 117dB

Result	Detector	Duration [ms]	1000	500	200	100	50	20	10	5	2	1	0.5	0.25
MAX	Fast	Indication [dB]	117.0	116.9	116.0	114.4	112.2	109.7	105.8	102.9	99.9	95.9	92.9	89.9
		Error [dB]	0.0	0.0	0.0	0.0	-0.0	-0.0	-0.1	0.0	-0.0	-0.1	-0.1	-0.1
	Slow	Indication [dB]	115.0	112.9	109.6	106.8	103.9	100.0	97.0	94.0	90.0	-	-	-
		Error [dB]	0.0	0.1	-0.0	-0.0	-0.0	-0.0	-0.0	-0.0	-0.0	-	-	-
SEL	-	Indication [dB]	117.0	114.0	110.0	107.0	104.0	100.0	97.0	94.0	90.0	86.9	83.9	80.9
		Error [dB]	0.0	-0.0	-0.0	0.0	-0.0	-0.0	0.0	-0.0	-0.0	-0.0	-0.1	-0.1

Issued to:

Fehily Timoney
J5 Plaza
North Park Business Park
North Road
Dublin 11

Calibration Reference

SLM200094

Test Date: 03/06/2020

Procedure: TP-SLM-1

Equipment

Item Calibrated:	Sound Level Meter	Model	977
Make:	Svantek	Serial Number:	69558

Calibration Procedure

The sound level meter was allowed to stabilize for a suitable period, as described in the manufacturer's instruction manual, in laboratory conditions. The sound level meter was calibrated by carrying out the verification tests detailed in IEC 61672-3 (2006), Periodic tests, specification of sound level meters. Tolerances for verification procedures are specified in IEC 61672-1 (2003).

Calibration Standards

Description	Serial Number
National Instruments PXI-4461	19C91D2
Stanford Research DS360	123803

The standards used in this calibration are traceable to NIST and/or other National Measurement Institutes (NMI's) that are signatories of the International Committee of Weights and Measures (CIPM) mutual recognition agreement (MRA).

Signed on behalf of Sonitus Systems:



Equipment Description

Model: Svantek **Serial Number:** 69558
Model: 977 **Microphone Model:** ACO 7052E

Ambient Conditions

Measurement conditions were within the tolerances defined in IEC 61672-1 and IEC 60942.

Barometric Pressure: 1030 hPa
Temperature: 23.4 °C
Relative Humidity: 39 %

Results Summary

IEC 61672 Test #	Test Description	Result
10	Self-generated noise	-
11	Frequency weighting (acoustical)	PASS
12	Frequency weighting (electrical)	PASS
13	Frequency and time weighting (1kHz)	PASS
14	Level linearity on reference level range	PASS
15	Level linearity with level range control	-
16	Toneburst response	PASS
17	Peak C sound level	PASS
18	Overload indication	PASS

As public evidence was available, from a testing organization responsible for approving the results of pattern evaluation tests, to demonstrate that the model of sound level meter fully conformed to the requirements for pattern evaluation described in IEC 61672:2003, the sound level meter tested is considered to conform to all the Class 1 requirements of IEC 61672:2003.

The manufacturer's guidelines concerning appropriate set up for measurement under various conditions should be observed during usage.

Prior to carrying out the verification tests the sound level meter was adjusted to read correctly using the acoustic calibrator held by the testing lab (Cirrus CR511ES, Serial number: 60871). The calibration procedure is described in the manufacturer's instruction manual.

Self-generated noise - IEC 61672-3 Test #10

SLM Measuring Mode: Leq

SLM Configuration	Freq. Weighting Network	SLM Reading
Microphone Installed	A	18.8
Microphone replaced by electrical input device fitted with short circuit	A	10.6
	C	10.6
	Z	10.6

Acoustical signal test of a frequency weighting - IEC 61672-3 Test #11

Range: reference level range

Frequency Weighting: C

Time Weighting: Slow

Input	Freq	Expected Level	Deviation	Tol +/-
94 dB	1000 Hz	94.0	0.0	1.0
	125 Hz	93.7	0.2	1.0
	4000 Hz	92.3	0.5	1.0

The frequency response was tested using an electrostatic actuator. Appropriate correction factors were applied where available from the manufacturer's instruction manual.

Electrical tests of frequency weighting - IEC 61672-3 Test #12

Range: reference level range

A-weighting

Freq	Expected Level	SLM Reading	Deviation	Tol +	Tol -
63	95.0	95.0	0.0	1.5	-1.5
125	95.0	94.9	-0.1	1.5	-1.5
250	95.0	94.9	-0.1	1.4	-1.4
500	95.0	95.0	0.0	1.4	-1.4
1000	95.0	95.0	0.0	1.1	-1.1
2000	95.0	95.0	0.0	1.6	-1.6
4000	95.0	95.1	0.1	1.6	-1.6
8000	95.0	95.1	0.1	2.1	-3.1
16000	95.0	94.7	-0.3	3.5	-17.0

C-weighting

Freq	Expected Level	SLM Reading	Deviation	Tol +	Tol -
63	95.0	95.0	0.0	1.5	-1.5
125	95.0	95.4	0.4	1.5	-1.5
250	95.0	95.0	0.0	1.4	-1.4
500	95.0	95.0	0.0	1.4	-1.4
1000	95.0	95.0	0.0	1.1	-1.1
2000	95.0	95.1	0.1	1.6	-1.6
4000	95.0	95.1	0.1	1.6	-1.6
8000	95.0	95.1	0.1	2.1	-3.1
16000	95.0	94.7	-0.3	3.5	-17.0

Linear

Freq	Expected Level	SLM Reading	Deviation	Tol +	Tol -
63	95.0	95.0	0.0	1.5	-1.5
125	95.0	95.0	0.0	1.5	-1.5
250	95.0	95.0	0.0	1.4	-1.4
500	95.0	95.0	0.0	1.4	-1.4
1000	95.0	95.0	0.0	1.1	-1.1
2000	95.0	95.0	0.0	1.6	-1.6
4000	95.0	95.0	0.0	1.6	-1.6
8000	95.0	95.0	0.0	2.1	-3.1
16000	95.0	95.0	0.0	3.5	-17.0

Frequency and Time Weightings at 1 kHz IEC 61672-3 Test #13

Range: reference level range

Time Weighting	Freq. Weighting	Expected Level	Deviation	Tol +/-
Fast	A	94.0	ref	
	C	94.0	0.0	0.2
Slow	A	94.0	0.0	0.2
LEQ	A	94.0	0.0	0.2

Linearity level on reference range - IEC 61672-3 Test #14

Input frequency: 8 kHz

SLM Measuring Mode: SPL

Range	Expected Level	SLM Reading	Deviation	Tol +/-
123 dB	94.0	94.0	0.0	1.1
	99.0	99.0	0.0	1.1
	104.0	104.0	0.0	1.1
	109.0	109.0	0.0	1.1
	114.0	114.0	0.0	1.1
	119.0	119.0	0.0	1.1
	124.0	124.0	0.0	1.1
	129.0	129.0	0.0	1.1
	134.0	134.0	0.0	1.1
	135.0	135.0	0.0	1.1
	136.0	136.0	0.0	1.1
	137.0	137.0	0.0	1.1
	89.0	89.0	0.0	1.1
	84.0	84.0	0.0	1.1
	79.0	79.0	0.0	1.1
	74.0	74.0	0.0	1.1
	69.0	69.0	0.0	1.1
	64.0	64.0	0.0	1.1
	59.0	59.0	0.0	1.1
	54.0	54.0	0.0	1.1
	49.0	49.1	0.1	1.1
	44.0	44.1	0.1	1.1
	39.0	39.1	0.1	1.1
	38.0	38.1	0.1	1.1
	37.0	37.2	0.2	1.1
	36.0	36.2	0.2	1.1
	35.0	35.3	0.3	1.1

Toneburst response - IEC 61672-3 Test #16

Range: reference level range

Burst Type	Response	Expected Level	SLM Reading	Deviation	Tol +	Tol -
0.25 ms	LAF _{MAX}	111.0	110.8	-0.2	0.8	-0.8
2.0 ms	LAF _{MAX}	120.0	119.9	-0.1	1.3	-1.3
200 ms	LAF _{MAX}	137.0	137.0	0.0	1.3	-3.3
2.0 ms	LAS _{MAX}	111.0	111.3	0.3	0.8	-0.8
200 ms	LAS _{MAX}	130.6	130.6	0.0	1.3	-3.3

Peak C sound level - IEC 61672-3 Test #17

Range: reference level range

Pulse Type	Freq	Expected Level	SLM Reading	Deviation	Tol +/-
1 cycle	8 kHz	135.4	135.3	-0.1	2.4
Pos ½ cycle	500 Hz	137.4	137.4	0	1.4
Neg ½ cycle	500 Hz	137.4	137.4	0	1.4

Overload indication IEC 61672-3 Test #18

Test Description	Overload at	Meas. Diff. (Pos – Neg)	Tol +/-
Pos. ½ cycle at 4 kHz	142.5		
Neg. ½ cycle at 4 kHz	142.5		
Level difference		0.0	1.8

Calibration Notes

1. The manufacturer's instruction manual was accessed through the manufacturer's website.
2. The sound level meter was powered by a regulated 9V power supply provided by the testing laboratory.

ENVIRONMENTAL CONDITIONS

Temperature	Relative humidity	Ambient pressure
29 °C	37%	1005 hPa

TEST EQUIPMENT

Item	Manufacturer	Model	Serial no.	Description
1.	SVANTEK	SVAN 401	127	Signal generator
2.	SVANTEK	SVAN 912A	9537	Sound & Vibration Analyser
3.	KEITHLEY	2000	0910165	Digital multimeter
4.	SVANTEK	SV33	48876	Acoustic calibrator
5.	SVANTEK	ST02	-	Microphone equivalent electrical impedance (18pF)
6.	DYTRAN	3333A	1376	Reference accelerometer

CONFORMITY & TEST DECLARATION

1. Hereby with SvanteK company declares that this instrument has been calibrated and tested in compliance with the internal ISO9001 procedures and meets all specification given in the Manual(s) or respectively surpass them.
2. The acoustic calibration was performed using the Sound Calibrator and is traceable to the GUM (Central Office of Measures) reference standard - sound level calibrator type 4231 No 2292773.
3. The vibrational calibration was performed using the Back-to-Back Comparison method and is traceable to the GUM (Central Office of Measures) reference standard - accelerometer type 8305 No 1435233.
4. The information appearing on this sheet has been compiled specifically for this instrument. This form is produced with advanced equipment & procedures which permit comprehensive quality assurance verification of all data supplied herein.
5. This calibration sheet shall not be reproduced except in full, without written permission of the SVANTEK Ltd.

Calibration specialist: Krzysztof Kubeł

Test date: 2018-06-01



ISO9001 certified

FACTORY CALIBRATION DATA OF THE SVAN 977A No. 69556

with preamplifier SVANTEK type SV12L No. 72145 and microphone ACO PACIFIC type 7052E No. 69608

SOUND LEVEL METER

1. CALIBRATION (electrical)

LEVEL METER function; Characteristic: A; $f_{ref}=1$ kHz; Input signal ≈ 110.9 dB;

Range	Low (120dB)	High (137dB)
Indication [dB]	114.0	114.0
Error [dB]	0.0	0.0

2. CALIBRATION* (acoustical)

LEVEL METER function; Range: High; Reference frequency: 1000 Hz; Sound Pressure Level: 113.99 dB.

Characteristic	Correct value [dB]	Indication [dB]	Error [dB]
Z	113.99	114.02	0.03
A	113.99	114.02	0.03
C	113.99	114.02	0.03

Calibration measured with the microphone ACO PACIFIC type 7052E No. 69608. Calibration factor: 0.63 dB.

3. LINEARITY TEST* (electrical)

LEVEL METER function; Range: Low; Characteristic: A; $f_{ref}=31.5$ Hz

Nominal result LEQ [dB]	24.0	25.0	26.0	28.0	30.0	40.0	60.0	80.0
Error [dB]	0.1	0.1	0.1	0.0	0.0	0.0	0.0	0.0

LEVEL METER function; Range: Low; Characteristic: A; $f_{ref}=1000$ Hz

Nominal result LEQ [dB]	24.0	25.0	26.0	28.0	30.0	40.0	60.0	80.0	100.0	120.0
Error [dB]	0.1	0.1	0.0	0.0	-0.0	-0.0	-0.0	-0.0	-0.0	-0.0

LEVEL METER function; Range: Low; Characteristic: A; $f_{ref}=8000$ Hz

Nominal result LEQ [dB]	24.0	25.0	26.0	28.0	30.0	40.0	60.0	80.0	100.0	119.0
Error [dB]	0.1	0.0	0.0	-0.0	-0.0	-0.0	-0.0	-0.0	0.0	-0.0

LEVEL METER function; Range: High; Characteristic: A; $f_{ref}=31.5$ Hz

Nominal result LEQ [dB]	35.0	36.0	37.0	38.0	40.0	60.0	80.0	97.0
Error [dB]	0.1	0.1	0.0	0.0	0.0	-0.0	-0.0	0.0

LEVEL METER function; Range: High; Characteristic: A; $f_{ref}=1000$ Hz

Nominal result LEQ [dB]	35.0	36.0	37.0	38.0	40.0	60.0	80.0	100.0	120.0	137.0
Error [dB]	0.1	0.1	0.0	0.0	0.0	-0.0	-0.0	-0.0	-0.0	0.0

LEVEL METER function; Range: High; Characteristic: A; $f_{ref}=8000$ Hz

Nominal result LEQ [dB]	35.0	36.0	37.0	38.0	40.0	60.0	80.0	100.0	120.0	136.0
Error [dB]	0.1	0.1	0.0	0.1	0.0	-0.0	-0.0	0.0	-0.0	0.0

4. TONE BURST RESPONSE*

LEVEL METER function; Characteristic: A; $f_{ref}=4000$ Hz; Burst duration: 2 s

Range: Low; Steady level nominal result = 117dB

Result	Detector	Duration [ms]	1000	500	200	100	50	20	10	5	2	1	0.5	0.25
MAX	Fast	Indication [dB]	117.0	116.9	116.0	114.4	113.3	108.7	105.8	102.9	99.0	96.0	92.9	89.4
		Error [dB]	0.0	0.0	0.0	-0.0	-0.0	-0.0	-0.1	0.0	-0.0	-0.1	-0.1	-0.1
	Slow	Indication [dB]	115.0	113.0	109.6	106.8	103.9	100.0	97.0	94.0	90.0	-	-	-
		Error [dB]	0.0	0.0	-0.0	-0.0	-0.0	-0.0	-0.0	-0.0	-0.0	-	-	-
SEL	-	Indication [dB]	117.0	114.0	110.0	107.0	104.0	100.0	97.0	94.0	90.0	87.0	83.9	80.9
		Error [dB]	0.0	-0.0	-0.0	0.0	-0.0	0.0	-0.0	-0.0	-0.0	-0.1	-0.1	-0.1



NSAI

National Metrology Laboratory

Certificate of Calibration

Issued to Fehily Timoney & Company
J5 Plaza
North Park Business Park
North Road
Dublin 11

Attention of Maureen Marsden

Certificate Number	213366
Item Calibrated	Bruel & Kjaer Type 4231 Sound Level Calibrator
Serial Number	2169963
ID Number	FTC 02
Order Number	6925
Date Received	20 Aug 2021
NML Procedure Number	AP-NM-13

Method The above calibrator was allowed to stabilize for a suitable period in laboratory conditions. It was then calibrated by measuring the sound pressure level generated in its measuring cavity (half-inch configuration). The calibrator's operating frequency was also measured.

Calibration Standards Norsonic 1504A Calibration System incorporating:
Agilent 34401A Multimeter, No. 0736 [Cal due: 10 Jun 2022]
B & K 4134 Measuring Microphone, No. 0744 [Cal due: 31 Jan 2022]
B & K 4228 Pistonphone, No. 0740 [Cal due: 31 Jan 2022]

Calibrated by



David Fleming

Approved by



Paul Hetherington

Date of Calibration

06 Sep 2021

Date of Issue

07 Sep 2021



This certificate is consistent with Calibration and Measurement Capabilities (CMC's) that are included in Appendix C of the Mutual Recognition Arrangement (MRA) drawn up by the International Committee for Weights and Measures. Under the MRA, all participating Institutes recognize the validity of each other's calibration certificates and measurement reports for quantities, ranges and measurement uncertainties specified in Appendix C (for details see www.bipm.org)